



Cabidlanan, Nabunturan, Davao de Oro

Bid Number: B-25-0154  
Date: 05/27/2025  
Page: 1 of 5

Purchase Request Number: 25-2761

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
1	1.0	unit	Refrigerated Centrifuge <ul style="list-style-type: none"> <li>• Large Capacity Refrigerated Centrifuge</li> <li>• An electrical interlock system</li> <li>• With Seven different lamp colors can be selected as desired for each program number, to indicate when the centrifuge is running or stopped.</li> <li>• Five unique buzzer melodies.</li> <li>• The inner cup for blood bags is designed for ease of use.</li> <li>• Acceptable imbalance up to 50 grams.</li> <li>• Step centrifugal function is available.</li> <li>• This centrifuge is compatible with automatic umbilical cord blood isolation unit.</li> <li>• With Foot Switch for hands free lid opening</li> <li>• Technical specification:               <ul style="list-style-type: none"> <li>o Speed: Maximum: 4000rpm</li> <li>o Speed setting: digital display 10rpm steps</li> <li>o Acceleration/Deceleration: 10-level selectors</li> <li>o RCF setting:                   <ul style="list-style-type: none"> <li>Maximum: 5,240g</li> <li>Setting: in 10 x g steps</li> </ul> </li> <li>o Maximum capacity: 200-500 mL of 12 blood bags</li> <li>o 400/500mL blood bags are inserted in the inner cup that has a central separator                   <ul style="list-style-type: none"> <li>minimise tilting when removing blood bags from one side</li> </ul> </li> <li>o Central separator provides two discrete locations for loading blood bags</li> <li>o Temperature Force:                   <ul style="list-style-type: none"> <li>Setting range: -9oC to 40oC in 1oC increments</li> <li>Indication: -9oC to 43oC in 1oC increments</li> </ul> </li> </ul> </li> <li>• Control system               <ul style="list-style-type: none"> <li>o Inverter Microprocessor Control</li> <li>o Speed</li> <li>o Centrifugal force</li> <li>o Time</li> <li>o Temperature</li> <li>o g sec</li> <li>o acceleration/deceleration</li> <li>o 50 memory</li> </ul> </li> </ul>	3,850,000.00	3,850,000.00		
			SUB-TOTAL		3,850,000.00		

Purpose:	FOR THE USE OF BLOOD AND PUBLIC LABORATORY 2ND QTR
----------	--

Delivery: PGSO

Period of Delivery: 60 DAYS



Cabidlanan, Nabunturan, Davao de Oro

Bid Number: B-25-0154  
Date: 05/27/2025  
Page: 2 of 5

Purchase Request Number: 25-2761

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
			<ul style="list-style-type: none"> <li>• Integrator setting and indication</li> <li>o 1.00 – 9.99 x 109 g sec</li> <li>• Timer setting and indication</li> <li>o 1 sec to 99 hour 59 min 59 sec, in 1 sec increment setting</li> <li>• Error and user message display</li> <li>o Lid open</li> <li>o Imbalance</li> <li>o Over speed</li> <li>o Abnormally high temperature</li> <li>o Function for detecting an occurrence of electrical abnormality in motor, inverter, speed sensor and temperature sensor</li> <li>• Footswitch for hands-free lid opening</li> <li>• Refrigerant: R4O4A</li> <li>• Power Supply: 220 – 240 Vac With AVR</li> </ul> <p>Standard Accessories:</p> <ul style="list-style-type: none"> <li>• Swinging bucket rotor</li> <li>• Inner cup:</li> <li>o 6 pcs</li> <li>o Has central separator</li> <li>• Footswitch</li> </ul> <p>Standard Requirements:</p> <ul style="list-style-type: none"> <li>• Complies with GMP Requirements</li> <li>• Compatible with automatic umbilical cord blood isolation unit</li> <li>• Acceptable imbalance of up to 50 grams</li> <li>• At least 7 operating status lamp colors selection</li> <li>• At least 5 buzzer melodies selection</li> <li>• Employed electrical interlock system</li> </ul> <p>Terms and Conditions</p> <p>1. General Provisions</p> <p>1.1 These terms and conditions ("Terms") govern the use, maintenance, and repair of laboratory equipment ("Equipment") both during the warranty period and after the expiration of the initial warranty.</p> <p>1.2 By using the Equipment, the provincial government of Davao de Oro ("Customer") agrees to comply with</p>				
			GRAND TOTAL		3,850,000.00		
Purpose: FOR THE USE OF BLOOD AND PUBLIC LABORATORY 2ND QTR Delivery: PGSO Period of Delivery: 60 DAYS							



Cabidlanan, Nabunturan, Davao de Oro

Bid Number: B-25-0154  
Date: 05/27/2025  
Page: 3 of 5

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
			<p>these Terms.</p> <p>2. Supplier's Responsibilities During Warranty Period</p> <p>2.1 The Supplier/Distributor shall provide a warranty for the Equipment for one year starting from the date of purchase.</p> <p>2.2 The Supplier/Distributor shall repair or replace defective parts caused by defects in materials or workmanship under normal use and maintenance at no cost to the Customer.</p> <p>2.3 The Supplier/Distributor is responsible for processing warranty claims upon receipt of proof of purchase and a detailed description of the defect from the Customer.</p> <p>2.4 During the warranty period, the Supplier/Distributor shall conduct regular monthly visits to check the Equipment's status and ensure its proper functioning.</p> <p>2.5 The Supplier/Distributor is required to have a service center within the Davao Region to ensure immediate response in case of Equipment breakdown.</p> <p>3. Warranty Exclusions</p> <p>3.1 The Supplier/Distributor is not responsible for the following: a. Damage caused by improper use, neglect, or unauthorized modifications by the Customer.</p> <p>b. Routine maintenance and consumable items.</p> <p>c. Damage resulting from external factors.</p> <p>3.2 The Supplier/Distributor will void the warranty if repairs or alterations are made by unauthorized personnel.</p> <p>4. Maintenance During Warranty</p> <p>4.1 The Supplier/Distributor shall provide preventive maintenance guidelines to the Customer to help maintain the warranty's validity.</p> <p>4.2 Preventive maintenance services may be performed by the Supplier/Distributor or an authorized service provider during the warranty period.</p> <p>5. Supplier's Responsibilities Post-Warranty</p> <p>5.1 After the warranty period, the Supplier/Distributor shall continue to offer repair and maintenance services on a chargeable basis.</p>				
			GRAND TOTAL		3,850,000.00		
Purpose: FOR THE USE OF BLOOD AND PUBLIC LABORATORY 2ND QTR Delivery: PGSO Period of Delivery: 60 DAYS							



Cabidlanan, Nabunturan, Davao de Oro

Bid Number: B-25-0154  
Date: 05/27/2025  
Page: 4 of 5

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
			<p>5.2 Upon receiving repair requests, the Supplier/Distributor shall provide an estimate of repair costs and obtain the Customer's approval before proceeding with any work.</p> <p>6. Replacement Parts</p> <p>6.1 The Supplier/Distributor shall supply genuine replacement parts for the Equipment upon request.</p> <p>6.2 The Supplier/Distributor shall ensure timely shipment of replacement parts, with associated costs borne by the Customer.</p> <p>7. Service Contracts</p> <p>7.1 The Supplier/Distributor shall offer optional service contracts for ongoing maintenance and support.</p> <p>7.2 These service contracts shall include various tiers with different service levels and response times to suit the Customer's needs.</p> <p>8. Technical Support</p> <p>8.1 The Supplier/Distributor shall provide technical support via phone, email, or onsite visits. Support may be subject to additional service fees as applicable.</p> <p>8.2 The Supplier/Distributor shall ensure a 24/7 technical assistance service to address urgent issues.</p> <p>8.3 The Supplier/Distributor is responsible for maintaining records of support requests and resolutions for reference.</p> <p>9. Liability</p> <p>9.1 The Supplier/Distributor is not liable for damage or malfunctions caused by improper use, unauthorized modifications, or repairs conducted by unauthorized personnel.</p> <p>9.2 The Supplier/Distributor's liability is limited to the repair or replacement of defective parts as stipulated in any applicable service contract or warranty agreement.</p> <p>10. Termination of Service</p> <p>10.1 The Supplier/Distributor reserves the right to terminate service if the Equipment is deemed unrepairable or if the Customer fails to comply with these Terms.</p> <p>10.2 The Supplier/Distributor shall provide a formal notice to the Customer before terminating any services.</p>				
			GRAND TOTAL		3,850,000.00		

Purpose:	FOR THE USE OF BLOOD AND PUBLIC LABORATORY 2ND QTR
Delivery:	PGSO
Period of Delivery:	60 DAYS



Cabidlanan, Nabunturan, Davao de Oro

Bid Number: B-25-0154  
Date: 05/27/2025  
Page: 5 of 5

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
			<p>10.3 Upon termination, the Supplier/Distributor shall provide the Customer with a report detailing the reasons for termination and any recommendations.</p> <p>11. Amendments</p> <p>11.1 The Supplier/Distributor reserves the right to amend these Terms at any time.</p> <p>11.2 The Supplier/Distributor shall notify the Customer of any changes in writing, with a minimum of [e.g., 30 days] notice prior to implementation.</p> <p>12. Governing Law</p> <p>12.1 These Terms are governed by and construed in accordance with the laws of the Republic of the Philippines.</p> <p>12.2 Any disputes arising out of or in connection with these Terms shall be subject to the exclusive jurisdiction of the Regional Trial Court in Nabunturan, Davao de Oro.</p>				
			GRAND TOTAL		3,850,000.00		

Purpose: FOR THE USE OF BLOOD AND PUBLIC LABORATORY 2ND QTR  
Delivery: PGSO  
Period of Delivery: 60 DAYS