



Cabidlanan, Nabunturan, Davao de Oro

Bid Number: B-25-0155  
Date: 06/04/2025  
Page: 1 of 4

Purchase Request Number: 25-3813

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
1	1.0	unit/s	<p>Newborn Hearing Machine</p> <p>Item Description Otoacoustic Emission Test (OAE), Brand New, Branded</p> <p>TECHNICAL SPECIFICATIONS</p> <p>Main Features :</p> <ul style="list-style-type: none"> <li>• Screen newborn &amp; infants with minimal cooperation from patients</li> <li>• Digital display; touch pad/touch screen</li> <li>• Fast and accurate results; 10 seconds or less per ear</li> <li>• Test at key speech frequencies: 2,3,4 and 5 kHz or equivalent</li> <li>• Can be fitted with different ear tips sizes</li> <li>• With memory storage and printer interface</li> <li>• Patient information can be downloaded in PC</li> <li>• Power: 220-240V, 60 Hz charger and built-in rechargeable battery</li> <li>• Results shows PASS/REFER</li> <li>• Supplier should provide Device Certification Program issued by Newborn Hearing Screening Reference Center (NHSRC)</li> <li>• Provision of Yearly Calibration Verification of the machine during the warranty period conducted by the supplier/manufacturer Calibration should be on site!</li> </ul> <p>Standard Accessories: (per unit)</p> <ul style="list-style-type: none"> <li>•OAE probe set with wire</li> <li>•Spare OAE probe set</li> <li>• Provision of printer</li> <li>•Spare variety of ear tips packs (2 sets)</li> <li>• With hard plastic carrying casing</li> </ul> <p>Note: Machine must not use cellphones as monitoring equipment</p> <p>"Note: The supplier shall include among it's special provisions, printing of the Province of Davao de Oro official logo/ inventory stickers on a portion of the equipment which is conspicuous but will not affect function; water proof (Material of the Province of Davao de Oro official logo and inventory stickers is high-</p>	1,012,000.00	1,012,000.00		
			SUB-TOTAL		1,012,000.00		
Purpose: FOR THE USE OF DDOPH-MONTEVISTA ( Otoacoustic Emission Test (OAE)) Delivery: DDOPH-MONTEVISTA Period of Delivery: 20 DAYS							



Cabidlanan, Nabunturan, Davao de Oro

Bid Number: B-25-0155  
Date: 06/04/2025  
Page: 2 of 4

Purchase Request Number: 25-3813

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
			<p>quality laminated vinyl stickers, water-proof or washed-proof;  - Size for inventory stickers for big equipment is 3 x 3 inches and for Province of Davao de Oro official logo is 3 inches diameter and;  For small equipment is 2 x 2 inches for inventory sticker and 2 inches diameter for Province of Davao de Oro official logo"</p> <p>From the Bidder:  1. Certification that they will submit Original Copy of Operator/User Manual in English language;  2. Certification that they will submit the Original Copy of service/technical and maintenance manual in English language;  3. Original brochure or manual (from the manufacturer) that contains all the technical specifications required by the bidding documents in English language; attached in the bidding documents;  4. The bidder must provide a Service Unit which the end-user can use in case the equipment will be pulled-out for repair/maintenance within the warranty period;  5. Copy of the Distributorship Agreement between the manufacturer and the local distributor (bidder);  6. Certification that the bidder will provide the training for at least two (2) end-users and two (2) for the maintenance staff;  7. List of completed similar projects/installations (same brand) with copies of contract or purchase orders attached for verification purposes; The supplier must present actual unit for evaluation of end-user and TWG members  8. Certification that the bidder shall be responsible for the notification, transportation, delivery, installations and commissioning at no cost to the government;  9. Certification that the Bid offer is in Philippine peso to include taxes and duties, transportation to site, delivery, installation, and testing expenses on site (hospital) by the bidder;</p>				
			GRAND TOTAL		1,012,000.00		

Purpose: FOR THE USE OF DDOPH-MONTEVISTA ( Otoacoustic Emission Test (OAE))  
Delivery: DDOPH-MONTEVISTA  
Period of Delivery: 20 DAYS



Cabidlanan, Nabunturan, Davao de Oro

Bid Number: B-25-0155  
Date: 06/04/2025  
Page: 3 of 4

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
			<p>10. Two (2) years warranty for parts and services. Warranty period starts/commence after the installation &amp; commissioning, end users' training on operation and acceptance by the end-user/hospital authorized personnel;</p> <p>11. Certification of free preventive maintenance visits (includes check-up, calibration verification and functionality test &amp; reports); semi-annually within the warranty period.</p> <p>"12. The supplier must have a Davao Region-based engineers/technicians and/or must be ready to respond to the request for repair within forty-eight (48) hours. The bidder must submit the names of Davao Region-based engineers;</p> <p>- The supplier shall conduct the necessary corrective maintenance within fourteen (14) calendar days upon notification from the end-user regarding the equipment breakdown / defects;</p> <p>- The supplier shall have the responsibilities and accountabilities that upon evaluation of equipment defects / faults, the equipment are appropriately repaired or replaced and shall be in good working condition thereafter;</p> <p>- The number of days where the equipment is unusable due to equipment defects / faults shall be added to the warranty period;</p> <p>- The supplier must submit a signed service report on the corrective maintenance done to the hospital and copy furnished the PGSO."</p> <p>From the Manufacturer:</p> <p>1. Valid certificate of manufacturer's compliance with ISO 13485: 2016; Medical Devices: Quality Management System - Requirement for regulatory purposes. The bidder should provide Certification of Conformity;</p> <p>2. Certification that the brand and/or model of the product has been in the local market for at least ten (10) years;</p> <p>3. Certification that the bidder/supplier has the engineer or technician trained to conduct the corrective and preventive maintenance of the product. The service engineer or technician must be presently</p>				
			GRAND TOTAL		1,012,000.00		

Purpose: FOR THE USE OF DDOPH-MONTEVISTA ( Otoacoustic Emission Test (OAE))

Delivery: DDOPH-MONTEVISTA

Period of Delivery: 20 DAYS



Cabidlanan, Nabunturan, Davao de Oro

Bid Number: B-25-0155  
Date: 06/04/2025  
Page: 4 of 4

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
			employed by the bidder/supplier or authorized by the manufacturer; 4. Certification that the equipment is a brand new, unused and not a discontinued model and subjected in the product recall; 5. Certification that the manufacturer shall guarantee the spare parts shall be available after the warranty period; 6. Certificate of training of engineers or technician on the applications/operations. The certificate shall have the following information: a. Name of Trainee b. Name/brand/model of equipment c. Inclusive dates of training d. Name of the trainer, date and venue				
			GRAND TOTAL		1,012,000.00		

Purpose: FOR THE USE OF DDOPH-MONTEVISTA ( Otoacoustic Emission Test (OAE))

Delivery: DDOPH-MONTEVISTA

Period of Delivery: 20 DAYS