



25-C0922

PROVINCIAL GOVERNMENT OF DAVAO DE ORO

Cabidianan, Nabunturan, Davao de Oro

TECHNICAL SPECIFICATIONS FORM

Purchase Request Number: 25-C0922

Bid Number: B-25-0219
 Date: 11/12/2025
 Page: 1 of 3

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
1	1.0	UNIT	<p>Billi Light/ Phototherapy unit s/N</p> <p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> • Type: Mobile • Radiate resource should be blue LED • Total spectral irradiance on the effective surface area • Must be at least 56uW/cm²/nm • Wavelength range: 450-465 nm or higher • LED lamps with enclosure; lamps last at least 20,000 hours • Effective surface area : 50 x 25 cm or higher @ 35 cm from the light source of equivalent • Height adjustment: 1350-1650 mm (from light source to floor) or equivalent • LCD display the total time of therapy lamp unit tilt up to 0-60 to better • Power requirements: 220 V, 60 Hz • Standard Accessories (per unit): <p>TERMS AND CONDITION:</p> <ul style="list-style-type: none"> • Provision of functional AVR (servo-type) at least 25% higher than rated capacity with multiple outlet (at least four outlets) • Warranty Period of 1year. 	79,000.00	79,000.00	_____	_____
2	1.0	unit/s	<p>ECG Machine</p> <p>TECHNICAL SPECIFICATIONS:</p> <p>Main features:</p> <ul style="list-style-type: none"> • Touch screen, LCD/LED Display • Channel digital with interpretation • Simultaneous acquisition of 12-lead ECG signal • Manual/auto/arrhythmia analysis or better • Multiple printing modes • 80mm maximum size of thermal paper, 3-channel formal recording • High-resolution thermal printing system • Built-in USB or RS232 	715,000.00	715,000.00	_____	_____
				SUB-TOTAL	794,000.00		

Purpose: FOR THE USE OF DAVAO DE ORO PROVINCIAL HOSPITAL - LAAK (MEDICAL APPARATUS) 4TH QUARTER

Delivery: DAVAO DE ORO PROVINCIAL HOSPITAL - LAAK

Period of Delivery: 10 DAYS



25-C0922

PROVINCIAL GOVERNMENT OF DAVAO DE ORO

Cabidianan, Nabunturan, Davao de Oro

TECHNICAL SPECIFICATIONS FORM

Purchase Request Number: 25-C0922

Bid Number: B-25-0219
 Date: 11/12/2025
 Page: 2 of 3

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
3	1.0	unit/s	<ul style="list-style-type: none"> • Lead: standard 12 leads, lead change automatically • Operation modes: manual/auto • Filter: AC, EMG or better • Drift filter: anti-drift system • Sensitivity: auto, 2.5, 5, 10, 20 and 40mm/mV or equivalent • Paper speed: 25 and 50mm/s <p>TERMS AND CONDITIONS: -Warranty period of 1 year.</p> <p>TECHNICAL SPECIFICATIONS:</p> <p>MAIN FEATURES: intended for monitoring physiological parameters of pregnant women during antepartum examinations, labor and <ul style="list-style-type: none"> digital display, at least 12 inches LCD and multi-display modes touch pad-button control on better Provision of the following functions and measurement <ul style="list-style-type: none"> single-FHR/Dual-FHR (for twins) - FHR measurement range: 50-240 bpm or higher - Ultrasound Frequency: 1 MHz or equivalent TOCO <ul style="list-style-type: none"> -TOCO range: 0-100% - Zero Mode: Automatic/manual Fetal Movement mark (FM) ECG (for mother) -NIBP (auto & manual) (for mother) <ul style="list-style-type: none"> - NIBP CUF INFLATING RANGE: 0-270 mmHg or Higher Sp02 (for mother) temperature - Temperature range: maximum of 45 or higher <ul style="list-style-type: none"> provision of built in-in printer/ recorder with z-fol/roll Thermal paper provision of alarms (visual and audible); monitor displays alarm messages provision of accessories holder power supply input: 220-240 VAC, 60 Hz with built in rechargeable pack Standard Accessories: (per unit) ultrasound (US) traducer (both) </p>	605,000.00	605,000.00	_____	_____
				SUB-TOTAL	1,399,000.00		

Purpose: FOR THE USE OF DAVAO DE ORO PROVINCIAL HOSPITAL - LAAK (MEDICAL APPARATUS) 4TH QUARTER

Delivery: DAVAO DE ORO PROVINCIAL HOSPITAL - LAAK

Period of Delivery: 10 DAYS



25-C0922

PROVINCIAL GOVERNMENT OF DAVAO DE ORO

Cabidianan, Nabunturan, Davao de Oro

TECHNICAL SPECIFICATIONS FORM

Purchase Request Number: 25-C0922

Bid Number: B-25-0219
 Date: 11/12/2025
 Page: 3 of 3

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
			for single & twins) TOCO Transducers Belts -Remove event marker - ECG Cable -Sp02 sensor with extension -NIBP cuff with extension -Temperature Sensor -Power Cable TERMS AND CONDITION: -Appropriate AVR (25% higher than the equipment rated value), 220 V, 60 Hz, SERVO motor type; with over-current -Chest Electrodes (disposable) 1 pack (100 pieces, minimum) for ECG -Spare Thermal Recorder Paper, 5 pieces - With rolling stand, medical grade steel construction, lockable wheel and accessory basket -Warranty period of 1 year.				
				GRAND TOTAL		1,399,000.00	

Purpose: FOR THE USE OF DAVAO DE ORO PROVINCIAL HOSPITAL - LAAK (MEDICAL APPARATUS) 4TH QUARTER
 Delivery: DAVAO DE ORO PROVINCIAL HOSPITAL - LAAK
 Period of Delivery: 10 DAYS